

# Calibration Procedures

## 2100P Portable Turbidimeter

### Step 1

Calibration procedures have already been completed by water quality staff.

### Step 2

Record calibration information found taped to the meter onto field data sheet.

2100P Portable Turbidimeter Keypad